

Testing Certificate SCINTO 2" x 1.5"Device No.: **15060166**Detector : **SEP171**

DR Cs-137 (true value)	DR (current) max \pm 20%
200 nSv/h	195 nSv/h
1000 nSv/h	961 nSv/h
10 μ Sv/h	9,97 μ Sv/h
100 μ Sv/h	94,4 μ Sv/h
Overflow > 200 μ Sv/h	X

Alarm Thresholds:

DLT I	X	CPS I	X
DLT II	X	CPS II	X
DLT III	X	CPS III	X

Delivered Options :

<ul style="list-style-type: none">• RS 232 cable• Accu charger• External probe• Probe cable	<input type="checkbox"/>	Software: V 1.10 – HW V 5.5
	<input type="checkbox"/>	Type:
	<input type="checkbox"/>	
	<input type="checkbox"/>	

Signature:**Date:****Saphymo Italia Srl**Vico Chiuso Paggi, 4/11
16128 Genova (IT)
P.IVA 03839470105

17.01.2011