

impianto **Eurocorporation**

APPARECCHIO

Saphymo Scinto

| | | |
|-------------------|-----------|-----------------|
| | MATRICOLA | 15060166 |
| taratura Saphymo | | 17-gen-11 |
| taratura ACCREDIA | | 11-giu-13 |

| NUCLIDE | E keV | riferimento μSv | val riferimento $\mu\text{Sv/h}$ | lettura $\mu\text{Sv/h}$ | rif/lettura | lettura/rif |
|---------|-------|-------------------------------|-------------------------------------|-----------------------------|-------------|-------------|
| Am-241 | 59,5 | 2,16E-001 | 1,30E+001 | 1,16E+002 | 0,11 | 8,92 |
| Cs-137 | 662 | 7,41E-001 | 4,44E+001 | 4,40E+001 | 1,01 | 0,99 |
| Co-60 | 1250 | 6,40E-001 | 3,84E+001 | 2,21E+001 | 1,74 | 0,58 |

| E keV | rapp al cesio | lettura/rif |
|-------|---------------|-------------|
| 59,5 | 9,0 | 8,92 |
| 662 | 1,0 | 0,99 |
| 1250 | 0,6 | 0,58 |

